

Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

Tender No. 08(C)/2023

UNIVERSITY OF CHAKWAL



TENDER DOCUMENT

For

PURCHASE OF LAB EQUIPMENT'S

Issued To: _____

Project Manager (PMU)
Ph. No.: 0543-553736



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

Table of Contents

S.No	Page #
Tender Notice	3
1. Invitation To the Bid	4
2. Instructions to the Bidders	4
Terms and Conditions of the Tender	6
3. Definitions	6
4. Tender Eligibility	6
5. Examination of the Tender Document	6
6. Amendment of the Tender Document	6
7. Bid Currency	7
8. Validity Period of the Bid	7
9. Bid Security	7
10. Bid Preparation and Submission	7
11. Financial Proposal	8
12. Modification/Withdrawal of the Tender	9
13. Bid Opening	9
14. Preliminary Examination	10
15. Determination of Responsiveness of the Bid	10
16. Technical Evaluation Criteria	10
17. Financial Proposal Evaluation	12
18. Rejection and Acceptance of the Tender/Bid	12
19. Contacting the Procuring Agency	13
20. Announcement of Evaluation Report	13
21. Award of Contract	13
22. Letter Of Acceptance (LOA)	13
23. Payment of Performance Guarantee (PG)	13
24. Refund of Bid Security (BS)	13
25. Issuance of Supply Order or Signing the Contract	14
26. Redressal of grievances by the procuring agency	14
General Conditions of Contract / Supply Order	14
27. Delivery of Items	14
28. Liquidated Damages	14
29. Inspection and Tests	15
30. Release of Performance Guarantee (PG)	15



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of
Chakwal City Campus"

31.	Contract Amendment	15
32.	Termination for Default	15
33.	Blacklisting	15
34.	Force Majeure	16
35.	Termination for Insolvency	16
36.	Forfeiture of Performance Security	16
37.	Payment	16
38.	Warranty	17
39.	Specification and Schedule of Requirement	18
Annexure- A	Technical Proposal Covering Letter	23
Annexure- B	Check List	24
Annexure- C	Bid Form	25
Annexure- D	AFFIDAVIT/BIDDER'S UNDERTAKING ON THE STAMP PAPER	26
Annexure- E	Financial Proposal Form	27
	(to be attached with Financial Proposal)	27
Annexure- F	Price Schedule	Error! Bookmark not defined.
Annexure- G	Contract Agreement Form	33



Tender Notice

Sealed tenders/ bids are invited from the firms registered with Sales Tax and Income Tax Departments for following Tender on the basis of **Single Stage (Two Envelope Procedure)** in terms of Rule No. **38 2(a)** of the Punjab Procurement Rules 2014 (PPRA Rules 2014).

Tender No.	Tender Name	Budgetary Amount	QTY	Closing Time and Date	Opening Time and Date
08(C)/2023	Purchase of Lab Equipment's for Department of Pharmacy, Chemistry and Botany from Project titled "Construction of Building at University of Chakwal City Campus"	39.803 Million	Detail in Tender Document	02-01-2024 till 11:00AM	02-01-2024 at 11:30AM

- Tender Document will be immediately available after publishing of this Tender Notice under Rule No. 25(1).
- Tender / Bid Document can be obtained from the Office of Project Manager, University of Chakwal between 08:00 am to 4:00 pm from Monday to Friday by depositing the **Tender Fee of Rs.2,000/-** through Deposit Slip of BOP in A/C No. **6010211599100019, Main Recurring Account of UOC** in the Bank of Punjab. In case the Tender / Bid Documents are downloaded from www.uoc.edu.pk or www.ppra.punjab.gov.pk, **Original Deposit Slip of BOP** must be attached with the **Technical Bid**.
- In term of rule 27 of PPRA Punjab Rules 2014, the Bid Security, of **Rs.796,060/-** which is equal to 2% of Budgeted Cost, in the form of "CDR, Bank Draft, Pay Order." in favor of Treasurer, University of Chakwal required to be submitted with the **Financial Bid**, without which the offer shall be rejected being non-responsive. **Copy of Bid Security** must be attached with the **Technical Bid**.
- Bids without supporting documents, undertaking, valid documentary evidence, and bids not conforming to terms and conditions given in the Tender Document will be liable for rejection.
- No supporting document will be accepted, at all, after opening of the Technical Bids. Bids received after due date and time or bids without Bid Security in shape of CDR, Bank Draft, Pay Order or less than required amount or Bid Security in shape of Cheque/ Cross Cheque shall be summarily rejected.
- The University reserve the right to modify/withdraw/cancel the bids/tender at any time prior to award of Contract, without assigning any reason or any obligation to inform the Tenderer of the grounds for the Purchaser's action, and without thereby incurring any liability to the Tenderer and the decision of the Purchaser shall be final.
- For obtaining any further information or clarifications, please contact the officer named below:

Project Manager (PMU)
University of Chakwal
Ph#0543-553736



1. Invitation To the Bid

- 1.1. Bids/Tenders are invited for **Purchase of Lab Equipment's for Department of Pharmacy, Chemistry and Botany from Project titled "Construction of Building at University of Chakwal City Campus** "from Sales Tax and Income Tax Registered firms.

2. Instructions to the Bidders

- 2.1. Procurement will be made under Punjab Procurement Rules (PPRA), 2014.
- 2.2. It will be clearly understood that the Terms and Conditions mentioned in this document are intended to be strictly enforced.
- 2.3. Bidders must ensure that they submit all the required documents indicated in the Tender / Bidding Documents at the time of opening of Technical Bids and no request for submission of missing documents will be entertained after opening of the Technical Bids.
- 2.4. Bids without supporting documents, undertaking, valid documentary evidence, and bids not conforming to terms and conditions given in the Tender Document will be liable for rejection. Bids received after due date and time, and bids without Bid Security in shape of Demand Draft, Pay Order, CDR or less than required amount or Bid Security in shape of Cheque or Cross Cheque shall be rejected.
- 2.5. No supporting document will be accepted, at all, after opening of the Technical Bids.
- 2.6. The bidder must quote rates of each item in a lot/package, partial bidding in a lot is allowed.
- 2.7. **Mode of Advertisement:**

As per Rule 12(2) of PPRA Punjab Rules, 2014.

2.8. Type of Open Competitive Bidding

As per Rule No. 38(2) a, single stage two envelope procedure shall be followed with details given below:

- 2.8.1. The bid shall comprise of a single package containing two separate envelopes. Each envelope shall contain separately the financial proposal and the technical proposal;
- 2.8.2. The envelopes shall be marked as "FINANCIAL PROPOSAL" and "TECHNICAL PROPOSAL" in bold and legible letters to avoid confusion;
- 2.8.3. Initially, only the envelope marked "TECHNICAL PROPOSAL" shall be opened; The envelope marked as "FINANCIAL PROPOSAL" shall be retained in the custody of procuring agency without being opened;
- 2.8.4. The Purchaser shall evaluate the technical proposal under **PPRA Rule # 32** and in a manner prescribed in this document, without reference to the price and reject any proposal which does not conform to the specified requirements as listed in said sections. During the technical evaluation no amendments in the technical proposal shall be permitted;
- 2.8.5. The financial proposals of technically qualified bidders shall be opened publicly at a time, and venue announced and communicated to the bidders in advance;
- 2.8.6. After the evaluation and approval of the technical proposal the procuring agency, shall at a time within the bid validity period, publicly open the financial proposals of the technically accepted and qualified bids only. The financial proposal of bids found technically non-responsive shall be returned un-opened to the respective bidders.



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

2.9. **Bid Data Sheet (Information for the Bidders)**

1	Procuring Agency	University of Chakwal
2	Tender Number	08(C)2023
3	Name of Tender	Purchase of Lab Equipment's for Department of Pharmacy, Chemistry and Botany from Project titled "Construction of Building at University of Chakwal City Campus"
4	Tender Document available place	Office of the Project Manager(PMU) University of Chakwal (Main Campus), Main Talagang Road, Chakwal.
5	Cost of Tender Document	Rs.2000/- payable at BOP through Deposit Slip in A/C No. 6010211599100019, Main Recurring Account of University of Chakwal.
6	Bid Security	Amount mentioned in Clause No. 9 of the Tender Document in shape of Bank Guarantee, CDR, Pay Order, Bank Draft etc., in favor of "Treasurer, University of Chakwal, Chakwal"
7	Performance Guarantee	10% of Contract Value after issuance of Letter of Acceptance
8	Tender Addressed to	Convener, Purchase Committee, University of Chakwal.
9	Contact Number	Ph. 0543-553736
10	Due date, time and place of submission of Tender Document	02-01-2024 till 11:00 AM Office of the Project Manager (PMU), University of Chakwal (Main Campus), Main Talagang Road, Chakwal.
11	Date, time and place of Technical Bid Opening	02-01-2024 till 11:30 AM at Seminar Hall , Ground Floor, Engineering Building, University of Chakwal (Main Campus), Main Talagang Road, Chakwal.
12	Date, time and place of the Financial Proposals	Shall be intimated subsequently to Technically Qualified firms

Terms and Conditions of the Tender

3. Definitions

- 3.1. "Purchaser" means the procuring agency i.e. University of Chakwal.
- 3.2. "UOC" means University of Chakwal.
- 3.3. "Bidder/Tenderer" means the Firm/Company/Supplier/Distributor that may provide or provides the Goods and related services to any of the public sector organization under the contract and have registered for the relevant business thereof.
- 3.4. "Contract" means the agreement entered into between the Purchaser and the Contractor, in form of Supply Order or as recorded in the Contract Form signed by the parties, including all Schedules and Attachments thereto and all documents incorporated by reference therein.
- 3.5. "Contractor/The Successful Bidder" means the person whose Tender has been accepted and awarded letter of Acceptance followed by the Supply Order or Contract by the Purchaser.
- 3.6. "Contract Value" means that portion of the Contract Price adjusted to give effect to such additions or deductions as are provided for in the Contract which is properly apportionable to the Goods or Services in question.
- 3.7. "Goods" means equipment, IT equipment, machinery, and/or other materials which the Contractor is required to supply to the Purchaser under the Contract.
- 3.8. "Services" means installation, configuration, deployment, commissioning, testing, training, support, after sale service, etc. of Goods and other such obligations which the Contractor is required to provide to the Purchaser under the Contract.
- 3.9. "PPRA" means Punjab Procurement Regulatory Authority.
- 3.10. The quantity of items to be procure can increased /decreased depending upon the availability of budget.

4. Tender Eligibility

Eligible Bidder/Tenderer is one who:

- 4.1. has valid registration certificates for Income Tax and Sales Tax;
- 4.2. is an active Income Tax Payer;
- 4.3. has got the experience in supply of relevant items;
- 4.4. conforms to the clause of "Responsiveness of Bid" given in this tender document;
- 4.5. Has not been blacklisted.

5. Examination of the Tender Document

The bidder/Tenderer is expected to examine the Tender Document, including all terms and conditions.

6. Amendment of the Tender Document

- 6.1. The Purchase Committee of University of Chakwal at any stage prior to the deadline for submission of the Tender, at its own initiative or in response to a clarification requested by the Bidder(s), amend the Tender Document, on any account, for any reason. All



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

- amendment(s) shall be part of the Tender Document and binding on the Bidder(s).
- 6.2. The Purchaser shall notify the amendment(s) in writing to the prospective Tenderers/Bidders.
- 6.3. The Purchaser may, at its exclusive discretion, amend the Tender Document to extend the deadline for the submission of the Tender, in which case all rights and obligations of the Purchaser and the Tenderers previously subject to the deadline shall thereafter be subject to the deadline as extended.

7. Bid Currency

Bidder should quote price in **Pak Rupees only** and payments shall also be made in Pakistan Rupees only.

8. Validity Period of the Bid

- 8.1. Validity period of the bids shall be **180 days**.
- 8.2. In exceptional circumstances, the Purchase Committee of University of Chakwal may ask the Bidders for an extension of the period of validity. The request and the responses shall be made in writing. A bidder accepting the request will not be required nor permitted to modify its tender.

9. Bid Security

- 9.1. Bidder will submit Bid Security drawn in the name of **Treasurer, University of Chakwal** detail given below:

Item Name	Budgetary Amount	Amount of Bid Security
Purchase of Lab Equipment's for Department of Pharmacy, Chemistry and Botany from Project titled "Construction of Building at University of Chakwal City Campus"	Rs.39.803 Million	Rs.796,060 /-

- 9.2. Cheque or Cross Cheque shall not be accepted at all.
- 9.3. The amount submitted as Bid Security shall be refunded to the unsuccessful bidders after the decision for the award of the said tender.
- 9.4. The Bid Security of Successful Bidder(s) may be converted as part of the Performance Guarantee for successful execution of the work.
- 9.5. Subject to the award of contract, the Bid Security in form of CDR, Demand Draft, Pay Order shall be returned to successful bidder against submission of Performance Guarantee
- 9.6. If the Bid Security is found less than the required amount then the bid will be rejected irrespective of the rates and the stage of the bid process.
- 9.7. The Bid Security may be forfeited if a Bidder:
- Refuses to accept Letter of Acceptance of the Bid; or
 - Fails to furnish Performance Security.

10. Bid Preparation and Submission

- 10.1. The Tender shall be filed in / accompanied by the prescribed Forms, Annexes, Schedules, Drawings, Documents, Brochures, Literature, etc. which shall be completely filled in,



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

stamped and signed by the Tenderer or his Authorized Representative. In case of copies, photocopies may be attested.

- 10.2. The Bids/Tenders should be submitted in two parts (1) **Technical Proposal** and (2) **Financial Proposal**. The Technical Proposal and Financial Proposal shall be submitted in separate covers.

10.3. **Technical Proposal**

The Technical Proposal will enable the Purchase/Technical Committee to evaluate whether the bidder is technically competent and capable of executing the order and the specifications offered by the bidder meet the ones given in the Tender/Bid Documents. Only those bids which qualify in the technical stage will be eligible for the Financial Proposal opening. The Financial Proposals of bidders who failed in the Technical stage will not be opened.

- 10.4. The Technical Proposal form as given in the Bid/Tender Document shall be filled, signed and stamped in all pages. The Purchase Committee will not be responsible for the errors committed in the bids by the bidders.
- 10.5. The Technical Proposal should not strictly contain any Price/Cost indications as such otherwise the bids will be summarily rejected.
- 10.6. The Bidder should quote only one brand/model/make of each item.
- 10.7. **Detail and Order of Documents to be furnished with the Technical proposal (Envelope A):**

The documents attached with the Technical Bid must be signed and stamped by the Authorized Representative of the Bidder. The documents attached must be numbered and attached in the following order:

- 10.7.1. Covering letter (*Annexure-A*) duly signed and stamped by authorized representative along with copy of Bid Security.
- 10.7.2. Copy of Income Tax Registration Certificate (*Annexure-B*).
- 10.7.3. Copy of Sales Tax Registration Certificate (*Annexure-C*).
- 10.7.4. Detailed specification of items (*Annexure-D*).
- 10.7.5. At least two relevant Supply Orders/Contracts received in the last three years (*Annexure-E*).
- 10.7.6. Affidavit/Undertaking on Stamp Paper (*Annexure-F*).

11. **Financial Proposal**

- 11.1. The Financial Proposal of the bidder shall also include the price break up of taxes/duties. All taxes/duties as applicable shall be responsibility of the bidders.
- 11.2. The cost quoted by the bidder shall be kept firm and unchanged for a period specified in the Bid/Tender Documents from the date of opening of the bids. The bidder shall keep the price firm/unchanged during the period of Contract including during the period of extension of time if any.
- 11.3. The quoted price will be inclusive of all taxes, duties, levies, insurance, freight (transportation charges), etc.
- 11.4. The Bid is liable for rejection if Financial Proposal contains conditional offer.
- 11.5. **Details to be furnished with Financial proposal (Envelope B)**



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

11.5.1. Financial Proposal Covering Letter (**Annexure G**).

11.5.2. Financial Proposal Form duly filled, signed and stamped by the Bidder (**Annexure H**).

11.5.3. Bid Security of **Rs.796,060/-**

11.6. Sealing the Bid

11.6.1. The Technical Proposal shall be placed in a separate cover (**Envelope-A**) and sealed appropriately. The Technical Proposal cover shall be marked as "**Technical Proposal for Lab Equipment's for Electronics Engineering Department**" – Bid Ref.: due on: ____ The "FROM" address and "TO" address shall be written without fail.

11.6.2. The Financial Proposal (**Envelope-B**) shall contain rates of item quoted by the bidder. The Financial Proposal cover shall be marked as "**Financial Proposal for Purchase of Lab Equipment's**" – Bid Ref.: due on: ____ The "FROM" address and "TO" address shall be written without fail.

11.6.3. **Out Cover:** The Technical Proposal cover (Envelope-A) and Financial Proposal cover (Envelope B) shall then be put in a single outer envelope and sealed. The outer cover shall be marked as "**Bid Purchase of Lab Equipment's**"–Bid Ref.: due on _____. The "FROM" address and "TO" address shall be written without fail.

11.7. Mode of Submission of Bids

11.7.1. The Bid should be addressed to "Convener, Purchase Committee, University of Chakwal".

11.7.2. The Bids received after Due Date and Time or Unsealed or incomplete or submitted by Fax or Email will be summarily rejected.

12. Modification/Withdrawal of the Tender

12.1. The Bidder may, by written notice served on the Purchaser, modify or withdraw the Tender after submission of the Tender, prior to the deadline for submission of the Tender.

12.2. The Tender, withdrawn after the deadline for submission of the Tender and prior to the expiration of the period of the Tender validity, shall result in forfeiture of the Bid Security.

13. Bid Opening

13.1. Technical Proposal Opening

The bid's outer cover and Technical Proposal cover will be opened by the Purchase Committee of University of Chakwal on the date and time as specified in the Tender Notice /Bid Data Sheet. The bids will be opened in the presence of the bidders who choose to be present. A maximum of two representatives for each bidder would be allowed to attend the Bid Opening.

13.2. Suppression of facts and misleading information

13.2.1. During the bid evaluation, if any suppression or misrepresentation of information is brought to the notice of the Purchase Committee, the Committee shall have the right to reject the Bid and if it happens so after selection of the Bidder, the Purchase Committee may terminate the Contract or award of the Contract or further processing of the Bid as the case may be and that will be without any compensation to the Bidder and the Bid



Security/Performance Guarantee, as the case may be, shall be forfeited.

- 13.2.2. It is the Bidder's responsibility to prove the Bidder's requisite qualification, experience and capacity to undertake the project to the entire satisfaction of the Purchase Committee failing which the Bid may be rejected.

14. Preliminary Examination

- 14.1. The Procuring Agency shall examine the bids to determine whether they are complete, whether any computational errors have been made, whether the required documents have been furnished and properly signed, and whether the bids are generally in order.
- 14.2. In the financial bids (at the time of opening the financial proposal) the arithmetical errors shall be rectified on the following basis. If there is a discrepancy between the unit price and the total price that is obtained by multiplying the unit price and quantity, the unit price shall prevail, and the total price shall be corrected. If the Bidders/Suppliers do not accept the correction of the errors, its bid shall be rejected. If there is a discrepancy between words and figures, the amount in words shall prevail.

15. Determination of Responsiveness of the Bid

- 15.1. The Purchaser shall determine the substantial responsiveness of the Tender to the Tender Document, prior to the Tender evaluation, on the basis of the contents of the Tender itself without recourse to extrinsic evidence. A substantially responsive Bid is one which:
- 15.1.1. meets eligibility criteria for the Bidder / the Goods / the Services;
 - 15.1.2. meets all the mandatory requirements of the evaluation criteria;
 - 15.1.3. meets the Technical Specifications for the Goods / the Services;
 - 15.1.4. meets the delivery period / point for the Goods / the Services;
 - 15.1.5. is accompanied by the required Bid Security;
 - 15.1.6. is otherwise complete and generally in order;
 - 15.1.7. Conforms to all terms and conditions of the Tender Document, without material deviation or reservation;
 - 15.1.8. which offers one Brand/Model/Make for each item (which does not contain any option);
- 15.2. A material deviation or reservation is one which affects the scope, quality or performance of the Goods or limits the Purchaser's rights or the Bidder's obligations under the Contract.
- 15.3. The Tender determined as not substantially responsive shall not subsequently be made responsive by the Tenderer by correction or withdrawal of the material deviation or reservation. However, the Purchaser may waive off any minor non-conformity or inconsistency or informality or irregularity in the Tender.

16. Technical Evaluation Criteria

The Bidders who have duly complied with the Eligibility/Qualification and Evaluation will be eligible for further processing.

The Bids which do not conform to the Technical Specifications or Bid conditions or the Bids from the Bidders without adequate capabilities for supply and maintenance / warranty services will be rejected.

The Eligible/Technically Qualified Bidders will be considered for further evaluation.

Under PPRA Rules # 31, Technical Bids will be evaluated on the basis of following criteria and

Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

Financial Bids of only those bidders will be opened who have fulfilled the criteria:

Evaluation Criteria			Requirement
Sales Tax Registration			Mandatory
Income Tax Registration			Mandatory
Professional Tax Registration			Mandatory
Authorization certificate from sole manufacturer			Mandatory
Conformance to the required specification of items given in Schedule of Requirement			Mandatory
At least three relevant Supply Orders/Contracts received in the last five years			Mandatory
Affidavit /Bidder's Undertaking on stamp paper			Mandatory
S. No.	Item Name and Description	Marks	Maximum Marks
1	Past Performance/ Experience of the Bidder (Reg. with GST/NTN)	--	10
1.1	1 - 3 year experience	2	--
1.2	4 - 8 year experience	4	--
1.3	9 - 15 year experience	6	--
1.4	Above 15	10	--
2	Relevant Experience (Purchase orders, supply orders, completion certificates must be attached, otherwise no marks shall be awarded.	--	20
2.1	1 - 5 year experience	5	--
2.2	6 - 10 year experience	10	--
2.3	11 & above year experience	20	--
3	Letter of Satisfaction after sale service by the client who has purchased minimum 05 million Please attach satisfaction /appreciation letter. (In case of missing information, no marks may be awarded 02 marks for each satisfaction certificate	10	10
4	Financial Position/ Status	--	10
4.1	Last sales tax paid Form	4	--
4.2	Bank Certificate (satisfactory)	2	--
4.3	Statement Worth (Min 20 M)	4	--
5	Technical Evaluation of quoted items	--	50
5.1	As per Serial # 39 Specifications	40	
5.2	Delivery schedule as per need	2.5	
5.3	List of clients / where such items delivered	2.5	
5.3	Guarantee / Warranty (as the case may be)	05	
TOTAL		--	100

The bidder must provide Verifiable documentary proof against all the mandatory requirement along with the Technical Proposal and no document will be received or considered after opening of the Technical Proposal.

Demonstration (If necessary) of Multimedia Projector will be held for confirmation of specifications of the quoted model.

Conformance to the required specification of items given in Schedule of Requirement will be evaluated by the Technical Evaluation Committee.

The Purchase Committee and Technical Evaluation Committee may ask for physical Demonstration of any item given in the Tender Document for confirmation of specifications.

17. Financial Proposal Evaluation

- 17.1. Technically qualified/successful bidder(s)/Tenderer(s) shall be called for opening of the Financial Proposal(s). The Financial Proposals will be opened in the presence of the Bidders at the time and venue indicated by the Purchaser accordingly. The technically Eligible/Successful Bidder(s)/Tenderer(s) or their authorized representatives against each Lot shall be allowed to take part in the Financial Proposal(s) opening against their relevant Lot(s).
- 17.2. Financial Proposal evaluation will be conducted under the Punjab Procurement Rules, 2014. The Price evaluation will include all duties, taxes and expenses etc. In case of any exemption of duties and taxes made by the Government in favor of the Purchaser, the contractor shall be bound to adjust the same in the Financial Proposal.

18. Rejection and Acceptance of the Tender/Bid

- 18.1. The Purchaser shall have the right, at his exclusive discretion, to increase / decrease the quantity of any or all item(s), under PPRA Rules 2014 without any change in unit prices or other terms and conditions, accept a Tender, reject any or all tender(s), cancel / annul the Tendering process at any time prior to award of Contract, without assigning any reason or any obligation to inform the Tenderer of the grounds for the Purchaser's action, and without thereby incurring any liability to the Tenderer and the decision of the Purchaser shall be final.
- 18.2. **The Tender / bid shall be rejected if:**
- 18.2.1. It is substantially non-responsive; or
- 18.2.2. The bidder does not meet any of the mandatory criteria mentioned in Clause No. 15; or
- 18.2.3. It does not contain the documentary proof against any of the mandatory criteria mentioned in Clause No. 15; or
- 18.2.4. **The bid is incomplete, conditional, alternative, late; or**
- 18.2.5. the bidder does not attach Bid Security in Shape of Demand Draft, Pay Order, CDR; or
- 18.2.6. the bid security is not attached or it is less than the required amount; or
- 18.2.7. the Bidder submits more than one Bids against one Tender; or
- 18.2.8. the Bidder tries to influence the Purchase Committee / Contract award; or
- 18.2.9. the Bidder engages in corrupt or fraudulent practices in competing for the Contract award; or
- 18.2.10. there is any discrepancy between bidding documents and bidder's proposal i.e. any non-conformity or inconsistency or informality or irregularity in the submitted bid; or
- 18.2.11. the Bidder submits any financial conditions as part of its bid which are not in conformity with tender document.



19. Contacting the Procuring Agency

- 19.1. No Bidder shall contact the Purchase Committee of UOC on any matter relating to its bid, from the time of the bid opening to the time the Contract is awarded.
- 19.2. Any effort by a Bidder to influence the Procuring Agency in its decisions on bid evaluation, bid comparison, or Contract Award will disqualify the bidder and rejection of the bid. Canvassing by any Bidder at any stage of the Tender evaluation is strictly prohibited.

20. Announcement of Evaluation Report

- 19.1 The Procuring Agency shall announce the results of bid evaluation in the form of a report giving justification for acceptance or rejection of bids at least ten days prior to the award of Contract. The report shall be made available on PPRA website and all the bidders shall be informed of this.

21. Award of Contract

- 21.1. The Tender will be awarded to the Lowest Evaluated Bidder(s) (item wise/Lot wise) who have been declared Technically Qualified.

22. Letter Of Acceptance (LOA)

After acceptance of the Bids by the CPC, Letter of Acceptance (LOA) will be issued only to the Successful Bidder(s).

23. Payment of Performance Guarantee (PG)

- 23.1. The Successful Bidder(s) will be required to remit the Performance Guarantee equivalent to 10% of the value of the contract price. The PG should be paid by way of Cash Deposit Receipt (CDR), Pay Order, Demand Draft issued from scheduled bank of Pakistan in favor of "Treasurer University of Chakwal, Chakwal" as unconditional Guarantee.
- 23.2. The Performance Guarantee will be forfeited if the Successful Bidder withdraws the Bid during the period of Bid validity specified in the Bid Documents or if the Bidder(s) fails to sign the contract.

24. Refund of Bid Security (BS)

The Bid Security (BS) of the Successful Bidder may be adjusted towards Performance Guarantee payable by the firm. If the successful Bidder(s) submits Performance Guarantee for the stipulated value in full by way of Cash Deposit Receipt (CDR), Bank Draft or Pay Order the BS will be refunded. The BS of the unsuccessful Bidder will be refunded on the written request of the Bidder. The Bid Security of the successful bidder shall be released upon his request provided the bidder submits the Performance Guarantee in the shape of Cash Deposit Receipt (CDR), Bank Draft or Pay Order.



25. Issuance of Supply Order or Signing the Contract

- 25.1. The Purchase Committee shall issue Supply Order or sign a Contract with the Successful bidder who has submitted the Performance Guarantee.
- 25.2. The Successful Bidder will provide the **stamp paper of 0.25% of total order value** for issuance of Supply Order or Signing the Contract.

26. Redressal of grievances by the procuring agency

- 26.1. Any bidder feeling aggrieved by any act of the procuring agency after the submission of his bid may lodge a written complaint concerning his grievances not later than 10 days after the announcement of the bid evaluation report.
- 26.2. The committee shall investigate and decide upon the complaint within fifteen days of the receipt of the complaint.
- 26.3. Mere fact of lodging of a complaint shall not warrant suspension of the procurement process.
- 26.4. Any bidder not satisfied with the decision of the committee of the procuring agency may lodge an appeal in the relevant court of jurisdiction.

General Conditions of Contract / Supply Order

27. Delivery of Items

- 27.1. The Supplier will be responsible for **delivery of Items** of Purchase of IT Equipment and in case of items which requires installation, the Bidder will be responsible for installation of items at UOC, Main Campus, Talagang Road, and Chakwal.
- 27.2. Delivery Period will be **Sixty (60) days** counted *after issuance of Letter of Acceptance*.
- 27.3. After delivery of items, the Bidder must get the items inspected at UOC, Main Campus, Talagang Road, and Chakwal.
- 27.4. The supplier will bear all costs associated with the preparation, delivery and installation of the Items and the Purchaser will in no case be responsible or liable for those costs. The supplier will make such arrangements to ensure safe delivery of goods. Any damage sustained during transportation / delivery will be rectified by the supplier at his cost.

28. Liquidated Damages

- 28.1. When the supplier fails to deliver or install the goods or both within the time period specified in the contract, the Purchase Committee may, without prejudice to any other remedy it may have under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to **0.25% of the price of the delayed goods per day of delay**, maximum up to 10% of the price of total value of the contract.
- 28.2. The Successful Bidder will be responsible to provide the delivery, Delivery Challan and Bill within the delivery period in order to avoid Late Delivery Charges.



29. Inspection and Tests

- 29.1. The Inspection Committee of UOC shall inspect and test the Goods supplied, the Services provided, under the Contract/Supply Order, to verify their conformity to the Technical Specifications.
- 29.2. Inspection will be done at University of Chakwal.
- 29.3. Inspection Committee may verify the authenticity of items.
- 29.4. After the inspection or test if the Inspection Committee is of the opinion that items do not conform to the specification and the criteria mentioned above, the Inspection Committee may reject them, and the supplier shall either replace the rejected goods or make all alterations necessary to meet the requirements of the specifications free of cost to University of Chakwal.

30. Release of Performance Guarantee (PG)

The Performance Guarantee will be refunded to the Successful Bidder(s) after one year of completion of delivery and installation of Goods (Inspection).

31. Contract Amendment

- 31.1. The Purchaser may, at any time, by written notice served on the Contractor, alter, amend, omit, increase, decrease or otherwise change the nature, quality, quantity and scope, of all / any of the Goods / the Services / the Works, in whole or in part.
- 31.2. No variation in or modification in the Contract shall be made, except by written amendment signed by both the Purchaser and the Contractor.

32. Termination for Default

The Purchase Committee of UOC may, without prejudice to any other remedy for breach of Contract, by written notice of default sent to the tenderer, terminate this Contract in whole or in part if:

- 32.1. the bidder fails to provide services within the period(s) specified in the Contract, or within any extension thereof granted by the Procuring entity.
- 32.2. the successful bidder fails to deliver goods and services as per its technical specifications offered in the bid
- 32.3. the successful bidder fails to perform any other obligation(s) under the Contract.
- 32.4. the bidder, in the judgment of the Purchase Committee has engaged in corrupt or fraudulent practices in competing for or in executing the Contract.

33. Blacklisting

If the Contractor fails / delays in performance of any of the obligations, under the Contract / Letter of Acceptance, violates any of the provisions of the Contract / Letter of Acceptance, commits breach of any of the terms and conditions of the Contract / Letter of Acceptance or found to have engaged in corrupt or fraudulent practices in competing for the award of contract / Letter of Acceptance or during the execution of the contract / Letter of Acceptance, the Purchaser may without prejudice to any other right of action / remedy it may have, blacklist the Contractor, either indefinitely or for a stated period, for future tenders in public sector, as per mechanism provided in Punjab Procurement Rules, 2014.



34. Force Majeure

Majeure means an act of nature or an event beyond the control of the Supplier and not involving the Supplier's fault or negligence directly or indirectly purporting to mis-planning, mismanagement and /or lack of ore sight to handle the situation. Such events may include but are not restricted to acts of the Procuring Agency in its sovereign capacity, wars or evolutions, fires, floods, earthquakes, strikes, epidemics, quarantine restrictions and freight embargoes. If a Force Majeure situation arises, the Suppliers hall promptly notify the Procuring Agency in writing with sufficient and valid evidence of such condition and the cause there of the Committee constituted for Redressal of grievances, shall examine the pros and cons of the case and all reasonable alternative means for completion of purchase order under the Contract and shall submit its recommendations to the competent authority. However, unless otherwise directed by the Procuring Agency in writing, the Supplier shall continue toper form its obligations under the Contract as far as is reasonably practical and shall seek reasonable alternative means for performance not prevented by the Force Majeure event.

35. Termination for Insolvency

The Procuring Agency may at any time terminate the Contract by giving written notice of 30days' time to the Supplier if the Supplier becomes bankrupt or otherwise insolvent. In this event, termination shall be without compensation to the Supplier, provided that such termination shall not prejudice or affect any right of action or remedy which has accrued or shall accrue thereafter to the Parties.

36. Forfeiture of Performance Security

- 36.1. If the Contractor fails / delays in performance of any of the obligations, under the Contract / violates any of the provisions of the Contract / commits breach of any of the terms and conditions of the Contract / Letter of Acceptance, the Purchaser may, without prejudice to any other right of action / remedy it may have, forfeit Performance Security of the Contractor.
- 36.2. Failure to supply required items/services within the specified time period will invoke penalty as specified in this document.

37. Payment

- 37.1. 100% payment will be made after withholding applicable taxes by UOC to the Successful Bidder after delivery and installation of goods and on receipt of the following documents:
- i. Triplicate Original Delivery Challan.
 - ii. Triplicate Original Bill/Invoice.
 - iii. Triplicate General Sales Tax Invoice
 - iv. Inspection/Completion report



38. Warranty

- 38.1.** Warranty of Equipment, Apparatus and instrument shall be one year.
- 38.2.** The Supplier shall confirm that the goods supplied under this contract are new and unused. The Supplier shall further warrant that all goods supplied under this contract shall have no defect arising from design materials or workmanship or from any act or omission of the supplier that may arise under the normal use of the supplied goods in the conditions prevailing in the University.
- 38.3.** The Purchase Committee shall notify the supplier in writing of any claim arising under the warranty. Upon receipt of such notice, the supplier shall repair or replace the defective goods or parts within **15days**, inclusive of, where applicable, the cost of inland delivery of the repaired or replaced goods or parts from the port of entry to the final destination and their installation.



39. Specification and Schedule of Requirement

Lab Equipment's For Botany Engineering Department

S/No	Name of Equipment	Specifications	Required Brand	Qty.	Quoted Model
01	DEPAERTMENT OF BOTANY				
01(a)	Clinical Centrifuge	Clinical centrifuge with 12 place rotor, speed range: 300-4500rpm in 10rpm/100rpm (fast turning the SPEED dial) increments, spins 7 different sizes of tubes and vacutainers using supplied inserts and adapters; e.g. 8x 15ml, 12x 10ml and 12x 5ml tubes; Speed accuracy of ± 20 rpm, Rotor type 12-bore angle rotor/24-bore angle rotor, features include safety function Lid interlock, sound alert, monitor for lid open/close condition, over-speed protection, over temperature detection, automatic internal diagnosis, ideal for the separation of serum, plasma, urea and blood samples, Input power 300W, Input voltage(frequency) AC 100-240V(50-60Hz) or Equivalent	As per approval of Sample	2	
01(b)	Mitosis Slide Set	Germany or Equivalent	As per approval of Sample	4	
01(c)	Meiosis Slide Set	Germany or Equivalent	As per approval of Sample	4	
01(d)	Hot Plate Stir 620 HPS	Features like dial controls for stir and heat, an indicator light for power and a 10-1/2" X 10-1/2" platform, Temperature: 280c Temperature Accuracy: ± 1 C Plate Size: 186x186 or 8-inch Plate: Glass Ceramic Capacity: 22 Liters Power: 1050 watt Stir Option: 0-1600 Rpm or Equivalent	As per approval of Sample	2	
01(e)	Vortex Mixture	Speed Range, 230V: 0 – 2,850 rpm, 120V: 0 – 3,400 rpm; Operating Modes: Touch or continuous; Ambient Operating Range, $+4^{\circ}$ to 65° C; Orbit of 4.5 mm or better, Dimensions (W x D x H), 5.5 x 6.3 x 5.1 in/14 x 16 x 13 cm; Weight, 4.84 lb/2.2 kg; Electrical, 230V~, 50 Hz or 120V~, 60 Hz or Equivalent	As per approval of Sample	2	
01(f)	Leaf Area Meter	Measures area, length, width, and perimeter with test units of Millimeter, square millimeter and calculates shape factor and aspect ratio, Precision level $\pm 2\%$, High resolution of	As per approval of Sample	1	

Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

		0.1mm, No calibration required, Stores data up to 1000 Groups measurements, Easy to use, Light weight and self-contained instrument with built-in data logger and LCD display, Rechargeable battery Measures objects up to 15 mm thick, 150 mm wide and 36 cm long, Micro Integrated IC, LCD Display and USB port for data transfer or Equivalent			
01(g)	Electrical Conductivity meter	Conductivity Range: Ranges, 0 to 200 uS/cm, 0 to 2 mS/cm, 0 to 20 mS/cm, 0 to 200 mS/cm & 0 to 1000 mS/cm & 0-3000uS/cm, Resistivity: 0Ω.cm-200MΩ.cm Salinity: 0.0 ppt...100.0 ppt TDS: 0.1 mg/L-199.9 g/L Temperature ATC/MTC-5 to 105 °C Resolution: ±0.1 % FS 3 Accuracy: ±0.1 % FS Connector: SP13 Calibration: 1 point up to 9 liquids of preset standards, Data Storage: up to 150 points or Equivalent	As per approval of Sample	2	
01(h)	pH Meter	Standardization range +/-2 pH (-2.00 to 16.00 pH; -2.000 to 16.000 pH); Temp. compensation 0-100 degree C; Display 4 digit LED with automatic polarity; Polarizing current 10 micro ampere, Resolution 0.01 pH; 0.001 pH, Accuracy (@25°C/77°F) ±0.01 pH; ±0.002 pH, Calibration automatic and up to three points (five points), 5 standard (7 standard) buffers available (1.68†, 4.01 or 3.00, 6.86, 7.01, 9.18, 10.01, 12.45) and two custom, automatic, -5.0 to 100.0°C (23.0 to 212.0°F) (using integral temperature sensor), Electrode Diagnostics standard mode with probe condition, response time and out of calibration range	As per approval of Sample	1	
01(i)	TDS Meter	TDS Range-0.0 to 199.9 mg/L (ppm); 0 to 1999 mg/L (ppm); 0.00 to 19.99 g/L (ppt); TDS Resolution-0.1 mg/L (ppm); 1 mg/L (ppm); 0.01 g/L (ppt); TDS Accuracy (@25°C/77°F)-±1% FS (excluding probe error); TDS Conversion Factor manual, one-point through TDS knob; Temperature Compensation-manual from 0 to 50°C (32 to 122°F) with β = 2%/°C; Power Supply 1 x 9V alkaline batteries; Power Consumption approximately 100 hours of continuous use; Environment-0 to 50°C (32 to 122°F); RH max 95%;Dimensions-145 x 80 x 36 mm (5.7 x 3.1 x 1.4"); Weight-230 g (8.1 oz); Ordering Information-Supplied with HI76301D conductivity probe, HI70032 1382 mg/L (ppm) calibration solution sachet, battery, and	As per approval of Sample	1	



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

		instruction manual; Recommended Probe-HI76301D four ring conductivity probe with DIN connector and 1 m (3.3') cable (included), Data logging EC allows to store up to 1000 log records of data, Data sets include readings, GLP data, date and time or Equivalent			
01(j)	Digital Temperature Logger	Measurement Range: -35 to +80°C with accuracy level of $\pm 0.9^{\circ}\text{F}$, Memory: 16382-32000 readings, USB Interface with software for Set-up data management and Data downloading, 2 User-Programmable Alarm Thresholds, Bright Red, Green and Orange LED Indication, Replaceable Internal Lithium Battery, IP 67 Protection with the protective cap, Resolution: 0.5 °C or Equivalent	As per approval of Sample	2	

S/No	Name of Equipment	Specifications	Required Brand	Qty.	Quoted Model
01	DEPAERTMENT OF CHEMISTRY				
01(a)	Hot plate with digital display	Stainless Steel 304 / Mild Steel / Cast Iron, Temperature: Up to 300°C, 220 Volts 50Hz, magnetic stirrer with each plate, automatic sensors,	As per approval of Sample	3	
01(b)	Infrared spectrophotometer	Range: 400-4000/cm; Temperature: 10°C ~30°C, humidity: less than 70%, 520 x 435 x 255mm, Transmission sample holder (Standard), sample analysis without KBR, along with computer/laptop compatible for the software and hardware of instruments	As per approval of Sample	01	
01(c)	Potentiostat Galvanostat	Three electrodes, voltage + 10 V, max current + 100 mA, software control of working electrodes, Frequency range: 10 μ Hz~1MHz, Current control range: $\pm 2\text{A}$, Potential resolution: 10 μ V; current resolution 1pA	As per approval of Sample	01	
01(d)	Spectrophotometer (Dual beam)	Wavelength range: 190 to 1100 nm Wavelength accuracy: $\pm 0.3\text{ nm}$ Wavelength repeatability: $\pm 0.05\text{ nm}$ (UV-Vis), bandpass: 1.0 nm;	As per approval of Sample	01	
01(e)	Fluorometer	XRF analyzers; Frequency: 50/60 Hz; Operating power: 100-240 VAC, A/1.2A	As per approval of Sample	01	
01(f)	Metal Housed Heating Mantles with control	Temperature range: upto 450 degree Electrical Heating Mantles with Controllers; Capacity: 5L	As per approval of Sample	02	



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of
Chakwal City Campus"

S/No	Name of Equipment	Specifications	Required Brand	Qty.	Quoted Model
01	DEPAERTMENT OF PHARAMACY				
01(a)	Moisture Analyzer	Touch color display, Maximum temperature range upto 200°C, Moisture content readability up to 0.01%, Heating module Halogen/IR emitter, Multiple sets of data storage, protection glass, sample weight 1mg to 50g. The supplier must provide training for the users of the instruments at site, after installation and commissioning.	As per approval of Sample	5	
01(b)	Automatic Melting point, Boiling point, Pore Point, Slip Point and Cloud Point apparatus	Three capillaries for melting, Three capillaries for boiling, temperature range from 10 to 400°C, Temperature gradients 0.1 to 20°C/min, with PC (DDR RAM, Hard disk 1TB, 21 inch LED Monitor), Printer, Melting point software, Cloud and Pour Point Tester with three test chambers, CFC refrigeration system, Separate digital control system for each chamber, sample jar capacity in each compartment = up to 4). The supplier must provide training for the users of the instruments at site, after installation and commissioning.	As per approval of Sample	1	
01(c)	Semi-automated fully compliant dissolution test apparatus	6 to 14 baskets with one litter capacity, Colored touch screen display, Automatic smooth noise proof up and down system, in built printer, Automatic sample collector in test tubes, temperature range 0 – 50°C, With PC (DDR RAM, Hard disk 1TB, 21 inch LED Monitor). The supplier must provide training for the users of the instruments at site, after installation and commissioning.	As per approval of Sample	1	
01(d)	Atomic Absorption Spectrometer with flame mode	Sample compartment one, Lamp capacity upto 6, three different band widths (0.1 to 1nm), Wave length range 180 to 900nm with wavelength accuracy of $\pm 0.25\text{nm}$, baseline stability of $\leq 0.004\text{A}/30\text{min}$, 6-lamp turret auto alignment, fully automated scan and peak-picking, 17 lamps (Ca, Mg, Al, Na, B, Fe, Cu, Ti, Cr, Ni, Si, Zr, K, Co, Mn, Hf, Sn, Pb). 50 mm size burner, validated operating and security softwares, good quality cuvettes, with PC (DDR RAM, Hard disk 500GB, 21 inch	As per approval of Sample	1	



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

		LED Monitor), good quality printer, 10KVA UPS with 15 -30min backup, and filled gas cylinders. On-site installation, commissioning, training and IQ/OQ (one vendor support) by factory-trained Engineer is required. Comprehensive support for equipment for a period of 12 months which will consist of the following at no extra cost: The warranty shall commence only upon successful completion of the Acceptance Test or commissioning. The supplier must provide training for the users of the instruments at site, after installation and commissioning.			
--	--	--	--	--	--

Note: Partial Bidding is allowed.

Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

Annexure- A Technical Proposal Covering Letter

To

The Convener,
Purchase Committee,
University of Chakwal,
Chakwal.

Dear Sir,

We are hereby submitting our Proposal, which includes the Technical Proposal and the Financial Proposal sealed in two separate envelopes. We have attached the **Technical Bid Form, Check List, Detailed Specifications** and the required **supporting documents** along with our Technical Bid.

Yours sincerely,

Authorized Signature

(In full and initials)

Name and Designation of Signatory Name of Firm Address

PMU-UOC



Page | 23

Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

Annexure- B Check List

The bidder must attach this list along with the Bid

1.1. Packing of the Bid

S#	Description	YES/NO
1	Whether the Bid is submitted in Two Envelopes	
2	Technical Proposal <i>Envelope A</i>	
3	Financial Proposal <i>Envelope B</i>	
4	Whether both Technical and Financial Proposals are put into another Outer Envelope .	

1.2. SUBMISSION AND ARRANGEMENT OF SUPPORTING DOCUMENTS

The Bidder must provide all the Supporting Documents, number all the pages of supporting documents, provide the page information and arrange the documents in the following order:

Envelope A: Enclosures of Technical Proposal		Attached YES/NO	Page#
1	Covering Letter		
2	Copy of Bid Security		
3	Copy of Income Tax Registration Certificate		
4	Copy of Sales Tax Registration Certificate		
5	Affidavit/Undertaking on the Stamp Paper		
6	Specifications of quoted items on the Letter Head of the bidder		
7	Two Supply Orders for supply of relevant items received in the past		
8	The specified catalogues / brochures of items quoted by the bidder		
Note: All the above documents and any other supporting document must be numbered and page number must be mentioned in the column specified for the purpose.			
Total Number of pages attached with the Technical Bid			Pages
8	Tender Document duly signed and stamped each page by the bidder must be attached at the end of the Technical Bid but numbering is not required for this document.		

Envelope B :Enclosures of Financial Proposal		Attached YES/No	Page #
1	Financial Proposal Form duly filled, signed and stamped by the bidder		
2	Price Schedule Form duly filled, signed and stamped by the bidder		
3	Bid Security in shape of Bank Draft, Pay Order or CDR.		



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

Annexure- C

Bid Form

A) Profile of the Bidder:

S#	Particulars	BIDDER
1	Name of the Company	
2	Year of Incorporation	
3	Registered Office	
	Address	
	Office Telephone Number	
	Fax Number	
4	Contact Person	
	Name of Authorized Representative	
	Personal Telephone Number	
	Email Address	
5	Registration Detail	
	NTN Registration Number	
	GST Registration Number	

B) Bid Security (Please do not mention amount of CDR)

S#	Particulars	Please Furnish Details
1	Name of the Bank	
2	Instrument Number and Date	



Annexure- D

AFFIDAVIT/BIDDER'S UNDERTAKING ON THE STAMP PAPER

Ref: Tender No. 08(C)/2023

1. We have examined the Tender/Bid Document and we undertake to meet the requirements regarding supply of Items, warranty and services as required and are prescribed in the Tender Document.
2. It is certified that the information furnished here in and as per the document submitted is true and correct and nothing has been concealed or tampered with.
3. We have read the provisions of Tender/Bid Document and confirm that these are acceptable to us. We further declare that additional conditions, variations, deviations, if any, found in our response shall not be given effect to.
4. We agree to unconditionally accept all the terms and conditions set out in the Tender/Bid Document
5. We undertake, if our Bid is accepted, to supply the items within the delivery period mentioned in the Tender Document.
6. We understand that no document regarding evaluation criteria will be accepted after opening of the Technical Bids and we are bound to provide all the documentary proofs regarding evaluation criteria or any other supporting document at the time of opening of Technical Bids.
7. We agree that the Purchase Committee of University of Chakwal is not bound to accept the lowest or any of the bids received. We also agree that the Purchase Committee reserves the right in absolute sense to reject all the products/ services specified in the Bid Response without assigning any reason whatsoever under PPRA Rules 2014.
8. We also declare that our Company/Organization is not blacklisted by any of the Federal or Provincial Government in Pakistan.

[Name and Signatures of authorized Person along with stamp]

PMU-UOC



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

Annexure- E Financial Proposal Form
(to be attached with Financial Proposal)

To

The Convener,
Purchase Committee,
University of Chakwal,
Chakwal.

Dear Sir,

With Reference to your Tender No. 08(C)/2023 PURCHASE OF LAB EQUIPMENT'S FROM ADP FUNDED SCHEME "CONSTRUCTION OF BUILDING AT UNIVERSITY OF CHAKWAL CITY CAMPUS"

Please find attached our Financial Proposal for the sum of Rs. (insert amount in words and figures). This amount is inclusive of all taxes.

We have attached the Bid Security of amount Rs. 796,060/- (Rupees Seven Lacs Ninety-Six Thousand Sixty Rupees Only) having CDR, Demand Draft, Pay Order No. along with our Financial Bid.

Yours sincerely,
Authorized Signature

PMU-UOC



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

Annexure- F Price Schedule

(Please attach this page along with Financial Bid or quote rates on the Letter Head of the Bidder)

Lab Equipment's for Botany Department

S/No	Name of Equipment	Specifications	Qty.	Quoted Model	Unit Price Including all Taxes	Total Price Including all Taxes
01		DEPARTMENT OF BOTANY				
01(a)	Clinical Centrifuge	Clinical centrifuge with 12 place rotor, speed range: 300-4500rpm in 10rpm/100rpm (fast turning the SPEED dial) increments, spins 7 different sizes of tubes and vacutainers using supplied inserts and adapters; e.g. 8x 15ml, 12x 10ml and 12x 5ml tubes; Speed accuracy of ± 20 rpm, Rotor type 12-bore angle rotor/24-bore angle rotor, features include safety function Lid interlock, sound alert, monitor for lid open/close condition, over-speed protection, over temperature detection, automatic internal diagnosis, ideal for the separation of serum, plasma, urea and blood samples, Input power 300W, Input voltage(frequency) AC 100-240V(50-60Hz) or Equivalent	2			
01(b)	Mitosis Slide Set	Germany or Equivalent	4			
01(c)	Meiosis Slide Set	Germany or Equivalent	4			
01(d)	Hot Plate Stir 620 HPS	Features like dial controls for stir and heat, an indicator light for power and a 10-1/2" X 10-1/2" platform, Temperature: 280c Temperature Accuracy: ± 1 C Plate Size: 186x186 or 8-inch Plate: Glass Ceramic Capacity: 22 Liters Power: 1050 watt Stir Option: 0-1600 Rpm or Equivalent	2			
01(e)	Vortex Mixture	Speed Range, 230V: 0 – 2,850 rpm, 120V: 0 – 3,400 rpm; Operating Modes: Touch or continuous; Ambient Operating Range, $+4^{\circ}$ to 65° C; Orbit of 4.5 mm or better, Dimensions (W x D x H), 5.5 x 6.3 x 5.1 in/14 x 16 x 13 cm; Weight, 4.84 lb/2.2 kg; Electrical, 230V~, 50 Hz or 120V~, 60 Hz or Equivalent	2			



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

01(f)	Leaf Area Meter	Measures area, length, width, and perimeter with test units of Millimeter, square millimeter and calculates shape factor and aspect ratio, Precision level $\pm 2\%$, High resolution of 0.1mm, No calibration required, Stores data up to 1000 Groups measurements, Easy to use, Light weight and self-contained instrument with built-in data logger and LCD display, Rechargeable battery Measures objects up to 15 mm thick, 150 mm wide and 36 cm long, Micro Integrated IC, LCD Display and USB port for data transfer or Equivalent	1			
01(g)	Electrical Conductivity meter	Conductivity Range: Ranges, 0 to 200 uS/cm, 0 to 2 mS/cm, 0 to 20 mS/cm, 0 to 200 mS/cm & 0 to 1000 mS/cm & 0-3000uS/cm, Resistivity: $0\Omega\text{.cm}$ -200M $\Omega\text{.cm}$ Salinity: 0.0 ppt...100.0 ppt TDS: 0.1 mg/L-199.9 g/L Temperature ATC/MTC-5 to 105 °C Resolution: $\pm 0.1\%$ FS 3 Accuracy: $\pm 0.1\%$ FS Connector: SP13 Calibration: 1 point up to 9 liquids of preset standards, Data Storage: up to 150 points or Equivalent	2			
01(h)	pH Meter	Standardization range ± 2 pH (-2.00 to 16.00 pH; -2.000 to 16.000 pH); Temp. compensation 0-100 degree C; Display 4 digit LED with automatic polarity; Polarizing current 10 micro ampere, Resolution 0.01 pH; 0.001 pH, Accuracy (@25°C/77°F) ± 0.01 pH; ± 0.002 pH, Calibration automatic and up to three points (five points), 5 standard (7 standard) buffers available (1.68†, 4.01 or 3.00, 6.86, 7.01, 9.18, 10.01, 12.45) and two custom, automatic, -5.0 to 100.0°C (23.0 to 212.0°F) (using integral temperature sensor), Electrode Diagnostics standard mode with probe condition, response time and out of calibration range	1			
01(i)	TDS Meter	TDS Range-0.0 to 199.9 mg/L (ppm); 0 to 1999 mg/L (ppm); 0.00 to 19.99 g/L (ppt); TDS Resolution-0.1 mg/L (ppm); 1 mg/L (ppm); 0.01 g/L (ppt); TDS Accuracy (@25°C/77°F)- $\pm 1\%$ FS (excluding probe error); TDS Conversion Factor manual, one-point	1			

Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

		through TDS knob; Temperature Compensation-manual from 0 to 50°C (32 to 122°F) with $\beta = 2\%/^{\circ}\text{C}$; Power Supply 1 x 9V alkaline batteries; Power Consumption approximately 100 hours of continuous use; Environment-0 to 50°C (32 to 122°F); RH max 95%;Dimensions-145 x 80 x 36 mm (5.7 x 3.1 x 1.4"); Weight-230 g (8.1 oz); Ordering Information-Supplied with HI76301D conductivity probe, HI70032 1382 mg/L (ppm) calibration solution sachet, battery, and instruction manual; Recommended Probe-HI76301D four ring conductivity probe with DIN connector and 1 m (3.3') cable (included), Data logging EC allows to store up to 1000 log records of data, Data sets include readings, GLP data, date and time or Equivalent				
01(j)	Digital Temperature Logger	Measurement Range: -35 to +80°C with accuracy level of $\pm 0.9^{\circ}\text{F}$, Memory: 16382-32000 readings, USB Interface with software for Set-up data management and Data downloading, 2 User-Programmable Alarm Thresholds, Bright Red, Green and Orange LED Indication, Replaceable Internal Lithium Battery, IP 67 Protection with the protective cap, Resolution: 0.5 °C or Equivalent	2			
		Total Cost				
		Total Cost (in Million)				

S/No	Name of Equipment	Specifications	Qty.	Quoted Model	Unit Price Including all Taxes	Total Price Including all Taxes
01		DEPARTMENT OF CHEMISTRY				
01(a)	Hot plate with digital display	Stainless Steel 304 / Mild Steel / Cast Iron, Temperature: Up to 300°C, 220 Volts 50Hz, magnetic stirrer with each plate, automatic sensors,	3			
01(b)	Infrared spectrophotometer	Range: 400-4000/cm; Temperature: 10°C ~ 30°C , humidity: less than 70%, 520 x	01			

Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

		435 x 255mm, Transmission sample holder (Standard), sample analysis without KBR, along with computer/laptop compatible for the software and hardware of instruments				
01(c)	Potentiostat Galvanostat	Three electrodes, voltage + 10 V, max current + 100 mA, software control of working electrodes, Frequency range: 10μHz~1MHz, Current control range: ±2A, Potential resolution: 10μV; current resolution 1pA	01			
01(d)	Spectrophotometer (Dual beam)	Wavelength range: 190 to 1100 nm Wavelength accuracy: +/-0.3 nm Wavelength repeatability: +/- 0.05 nm (UV-Vis), bandpass: 1.0 nm;	01			
01(e)	Fluorometer	XRF analyzers; Frequency: 50/60 Hz; Operating power: 100-240 VAC, A/1.2A	01			
01(f)	Metal Housed Heating Mantles with control	Temperature range: upto 450 degree Electrical Heating Mantles with Controllers; Capacity: 5L	01			

S/No	Name of Equipment	Specifications	Qty.	Quoted Model	Unit Price Including all Taxes	Total Price Including all Taxes
01		DEPARTMENT OF PHARMACY				
01(a)	Moisture Analyzer	Touch color display, Maximum temperature range upto 200°C, Moisture content readability up to 0.01%, Heating module Halogen/IR emitter, Multiple sets of data storage, protection glass, sample weight 1mg to 50g	5			
01(b)	Automatic Melting point,	Three capillaries for melting, Three capillaries for boiling, temperature	1			



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

	Boiling point, Pore Point, Slip Point and Cloud Point apparatus	range from 10 to 400°C, Temperature gradients 0.1 to 20°C/min, with PC (DDR RAM, Hard disk 1TB, 21 inch LED Monitor), Printer, Melting point software, Cloud and Pour Point Tester with three test chambers, CFC refrigeration system, Separate digital control system for each chamber, sample jar capacity in each compartment = up to 4)				
01(c)	Semi-automated fully compliant dissolution test apparatus	6 to 14 baskets with one litter capacity, Colored touch screen display, Automatic smooth noise proof up and down system, in built printer, Automatic sample collector in test tubes, temperature range 0 – 50°C, With PC (DDR RAM, Hard disk 1TB, 21 inch LED Monitor)	1			
01(d)	Atomic Absorption Spectrometer with flame mode	three different band widths (0.1 to 1nm), Wave length range 180 to 900nm with wavelength accuracy of $\pm 0.25\text{nm}$, baseline stability of $\leq 0.004\text{A}/30\text{min}$, 6-lamp turret auto alignment, fully automated scan and peak-picking, 17 lamps (Ca, Mg, Al, Na, B, Fe, Cu, Ti, Cr, Ni, Si, Zr, K, Co, Mn, Hf, Sn, Pb). 50 mm size burner, validated operating and security softwares, good quality cuvettes, with PC (DDR RAM, Hard disk 500GB, 21 inch LED Monitor), good quality printer, 10KVA UPS with 15 -30min backup, and filled gas cylinders. On-site installation, commissioning, training and IQ/OQ (one vendor support) by factory-trained Engineer is required. Comprehensive support for equipment for a period of 12 months which will consist of the following at no extra cost: The warranty shall commence only upon successful completion of the Acceptance Test or commissioning. The supplier must provide training for the users of the instruments at site, after installation and commissioning.	1			



Purchase of Lab Equipment's from the Project Titled "Construction of Building at University of Chakwal City Campus"

Annexure- G Contract Agreement Form

THIS AGREEMENT made the [day] day of [month] [year] between [University of Chakwal] (hereinafter called "the University") of the one part and [name and address of Supplier] (hereinafter called "the Supplier") of the other part:

WHEREAS the University invited Tenders for certain goods and related services, viz, [brief description of goods and related services] and has accepted a Tender by the Supplier (Supply of IT Equipment) for the supply of those goods and related services in the sum of Rs. [Contract Price in figures and in words] (hereinafter called "the Contract Price").

NOW THEREFORE the parties hereby agree as follow:

1- The following documents attached shall be deemed to form and integral part of this Contract:

- i- Tender/Bid Document
- ii- Letter of Acceptance
- iii- Performance Guarantee equal to 10% of Contract Price in shape of CDR, Demand Draft or Pay Order.

2- The Terms and Conditions of Supply Order/Contract given in the Tender/Bid Document will be applicable.

3- The mutual rights and obligations of the University and the Supplier will be preserved in the light of the Terms and Conditions mentioned in the Tender /Bid Document.

IN WITNESS whereof the parties have caused this Contract to be executed in accordance with the laws of Pakistan on the day, month and year written above.

For University of Chakwal

For the Supplier:

Signature

Signature

Print Name

Title

PMU-UOC

